

December 16, 2019

Holly Fischer
DaVinci Academy of Arts and Science
532 Bunker Lake Blvd NE.
Ham Lake, MN 55304

RE: Project: Drinking Waters
Pace Project No.: 10501112

Dear Holly Fischer:

Enclosed are the analytical results for sample(s) received by the laboratory on December 02, 2019. The results relate only to the samples included in this report. Results reported herein conform to the most current, applicable TNI/NELAC standards and the laboratory's Quality Assurance Manual, where applicable, unless otherwise noted in the body of the report.

If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Jared Dickinson
jared.dickinson@pacelabs.com
(612)607-1700
Project Manager

Enclosures



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

CERTIFICATIONS

Project: Drinking Waters

Pace Project No.: 10501112

Pace Analytical Services Minneapolis

A2LA Certification #: 2926.01
Alabama Certification #: 40770
Alaska Contaminated Sites Certification #: 17-009
Alaska DW Certification #: MN00064
Arizona Certification #: AZ0014
Arkansas DW Certification #: MN00064
Arkansas WW Certification #: 88-0680
California Certification #: 2929
CNMI Saipan Certification #: MP0003
Colorado Certification #: MN00064
Connecticut Certification #: PH-0256
EPA Region 8+Wyoming DW Certification #: via MN 027-053-137
Florida Certification #: E87605
Georgia Certification #: 959
Guam EPA Certification #: MN00064
Hawaii Certification #: MN00064
Idaho Certification #: MN00064
Illinois Certification #: 200011
Indiana Certification #: C-MN-01
Iowa Certification #: 368
Kansas Certification #: E-10167
Kentucky DW Certification #: 90062
Kentucky WW Certification #: 90062
Louisiana DEQ Certification #: 03086
Louisiana DW Certification #: MN00064
Maine Certification #: MN00064
Maryland Certification #: 322
Massachusetts Certification #: M-MN064
Massachusetts DWP Certification #: via MN 027-053-137
Michigan Certification #: 9909
Minnesota Certification #: 027-053-137

Minnesota Dept of Ag Certification #: via MN 027-053-137
Minnesota Petrofund Certification #: 1240
Mississippi Certification #: MN00064
Missouri Certification #: 10100
Montana Certification #: CERT0092
Nebraska Certification #: NE-OS-18-06
Nevada Certification #: MN00064
New Hampshire Certification #: 2081
New Jersey Certification #: MN002
New York Certification #: 11647
North Carolina DW Certification #: 27700
North Carolina WW Certification #: 530
North Dakota Certification #: R-036
Ohio DW Certification #: 41244
Ohio VAP Certification #: CL101
Oklahoma Certification #: 9507
Oregon Primary Certification #: MN300001
Oregon Secondary Certification #: MN200001
Pennsylvania Certification #: 68-00563
Puerto Rico Certification #: MN00064
South Carolina Certification #: 74003001
Tennessee Certification #: TN02818
Texas Certification #: T104704192
Utah Certification #: MN00064
Vermont Certification #: VT-027053137
Virginia Certification #: 460163
Washington Certification #: C486
West Virginia DEP Certification #: 382
West Virginia DW Certification #: 9952 C
Wisconsin Certification #: 999407970
Wyoming UST Certification #: via A2LA 2926.01

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE SUMMARY

Project: Drinking Waters

Pace Project No.: 10501112

Lab ID	Sample ID	Matrix	Date Collected	Date Received
10501112001	7	Drinking Water	11/27/19 09:40	12/02/19 18:20
10501112002	8	Drinking Water	11/27/19 09:40	12/02/19 18:20
10501112003	9	Drinking Water	11/27/19 09:40	12/02/19 18:20
10501112004	10	Drinking Water	11/27/19 09:40	12/02/19 18:20
10501112005	11	Drinking Water	11/27/19 09:40	12/02/19 18:20
10501112006	12	Drinking Water	11/27/19 09:40	12/02/19 18:20
10501112007	13	Drinking Water	11/27/19 09:40	12/02/19 18:20
10501112008	14	Drinking Water	11/27/19 09:40	12/02/19 18:20
10501112009	15	Drinking Water	11/27/19 09:40	12/02/19 18:20
10501112010	16	Drinking Water	11/27/19 09:40	12/02/19 18:20
10501112011	17	Drinking Water	11/27/19 09:40	12/02/19 18:20
10501112012	18	Drinking Water	11/27/19 09:40	12/02/19 18:20
10501112013	38	Drinking Water	11/27/19 08:45	12/02/19 18:20
10501112014	39	Drinking Water	11/27/19 08:45	12/02/19 18:20
10501112015	40	Drinking Water	11/27/19 08:45	12/02/19 18:20
10501112016	41	Drinking Water	11/27/19 08:50	12/02/19 18:20
10501112017	42	Drinking Water	11/27/19 08:50	12/02/19 18:20
10501112018	43	Drinking Water	11/27/19 08:50	12/02/19 18:20
10501112019	44	Drinking Water	11/27/19 08:55	12/02/19 18:20
10501112020	45	Drinking Water	11/27/19 08:55	12/02/19 18:20
10501112021	46	Drinking Water	11/27/19 08:55	12/02/19 18:20
10501112022	29	Drinking Water	11/27/19 09:10	12/02/19 18:20
10501112023	30	Drinking Water	11/27/19 09:10	12/02/19 18:20
10501112024	31	Drinking Water	11/27/19 09:10	12/02/19 18:20
10501112025	32	Drinking Water	11/27/19 09:10	12/02/19 18:20
10501112026	33	Drinking Water	11/27/19 09:10	12/02/19 18:20
10501112027	34	Drinking Water	11/27/19 09:10	12/02/19 18:20
10501112028	35	Drinking Water	11/27/19 09:10	12/02/19 18:20
10501112029	36	Drinking Water	11/27/19 09:10	12/02/19 18:20
10501112030	37	Drinking Water	11/27/19 09:10	12/02/19 18:20
10501112031	1	Drinking Water	11/27/19 09:20	12/02/19 18:20
10501112032	2	Drinking Water	11/27/19 09:20	12/02/19 18:20
10501112033	3	Drinking Water	11/27/19 09:20	12/02/19 18:20
10501112034	4	Drinking Water	11/27/19 09:20	12/02/19 18:20
10501112035	5	Drinking Water	11/27/19 09:20	12/02/19 18:20
10501112036	6	Drinking Water	11/27/19 09:20	12/02/19 18:20
10501112037	19	Drinking Water	11/27/19 10:00	12/02/19 18:20

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE SUMMARY

Project: Drinking Waters

Pace Project No.: 10501112

Lab ID	Sample ID	Matrix	Date Collected	Date Received
10501112038	20	Drinking Water	11/27/19 10:00	12/02/19 18:20
10501112039	21	Drinking Water	11/27/19 10:00	12/02/19 18:20
10501112040	22	Drinking Water	11/27/19 10:00	12/02/19 18:20
10501112041	23	Drinking Water	11/27/19 10:00	12/02/19 18:20
10501112042	24	Drinking Water	11/27/19 10:00	12/02/19 18:20
10501112043	26	Drinking Water	11/27/19 10:00	12/02/19 18:20
10501112044	27	Drinking Water	11/27/19 10:00	12/02/19 18:20

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE ANALYTE COUNT

Project: Drinking Waters

Pace Project No.: 10501112

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
10501112001	7	EPA 200.8	BWB	1	PASI-M
10501112002	8	EPA 200.8	BWB	1	PASI-M
10501112003	9	EPA 200.8	BWB	1	PASI-M
10501112004	10	EPA 200.8	BWB	1	PASI-M
10501112005	11	EPA 200.8	BWB	1	PASI-M
10501112006	12	EPA 200.8	BWB	1	PASI-M
10501112007	13	EPA 200.8	BWB	1	PASI-M
10501112008	14	EPA 200.8	BWB	1	PASI-M
10501112009	15	EPA 200.8	BWB	1	PASI-M
10501112010	16	EPA 200.8	BWB	1	PASI-M
10501112011	17	EPA 200.8	BWB	1	PASI-M
10501112012	18	EPA 200.8	BWB	1	PASI-M
10501112013	38	EPA 200.8	BWB	1	PASI-M
10501112014	39	EPA 200.8	BWB	1	PASI-M
10501112015	40	EPA 200.8	BWB	1	PASI-M
10501112016	41	EPA 200.8	BWB	1	PASI-M
10501112017	42	EPA 200.8	BWB	1	PASI-M
10501112018	43	EPA 200.8	BWB	1	PASI-M
10501112019	44	EPA 200.8	BWB	1	PASI-M
10501112020	45	EPA 200.8	BWB	1	PASI-M
10501112021	46	EPA 200.8	BWB	1	PASI-M
10501112022	29	EPA 200.8	BWB	1	PASI-M
10501112023	30	EPA 200.8	BWB	1	PASI-M
10501112024	31	EPA 200.8	BWB	1	PASI-M
10501112025	32	EPA 200.8	BWB	1	PASI-M
10501112026	33	EPA 200.8	BWB	1	PASI-M
10501112027	34	EPA 200.8	BWB	1	PASI-M
10501112028	35	EPA 200.8	BWB	1	PASI-M
10501112029	36	EPA 200.8	BWB	1	PASI-M
10501112030	37	EPA 200.8	BWB	1	PASI-M
10501112031	1	EPA 200.8	BWB	1	PASI-M
10501112032	2	EPA 200.8	BWB	1	PASI-M
10501112033	3	EPA 200.8	BWB	1	PASI-M
10501112034	4	EPA 200.8	BWB	1	PASI-M
10501112035	5	EPA 200.8	BWB	1	PASI-M
10501112036	6	EPA 200.8	BWB	1	PASI-M
10501112037	19	EPA 200.8	BWB	1	PASI-M

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE ANALYTE COUNT

Project: Drinking Waters

Pace Project No.: 10501112

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
10501112038	20	EPA 200.8	BWB	1	PASI-M
10501112039	21	EPA 200.8	BWB	1	PASI-M
10501112040	22	EPA 200.8	BWB	1	PASI-M
10501112041	23	EPA 200.8	BWB	1	PASI-M
10501112042	24	EPA 200.8	BWB	1	PASI-M
10501112043	26	EPA 200.8	BWB	1	PASI-M
10501112044	27	EPA 200.8	BWB	1	PASI-M

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Drinking Waters

Pace Project No.: 10501112

Sample: 7		Lab ID: 10501112001	Collected: 11/27/19 09:40	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.5	ug/L	0.10	1		12/13/19 11:16	7439-92-1	
Sample: 8		Lab ID: 10501112002	Collected: 11/27/19 09:40	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.8	ug/L	0.10	1		12/13/19 11:27	7439-92-1	
Sample: 9		Lab ID: 10501112003	Collected: 11/27/19 09:40	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.90	ug/L	0.10	1		12/13/19 11:28	7439-92-1	
Sample: 10		Lab ID: 10501112004	Collected: 11/27/19 09:40	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.2	ug/L	0.10	1		12/13/19 11:30	7439-92-1	
Sample: 11		Lab ID: 10501112005	Collected: 11/27/19 09:40	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	ND	ug/L	0.10	1		12/13/19 11:32	7439-92-1	
Sample: 12		Lab ID: 10501112006	Collected: 11/27/19 09:40	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	ND	ug/L	0.10	1		12/13/19 11:34	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Drinking Waters

Pace Project No.: 10501112

Sample: 13		Lab ID: 10501112007	Collected: 11/27/19 09:40	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	ND	ug/L	0.10	1		12/13/19 11:36	7439-92-1	
Sample: 14		Lab ID: 10501112008	Collected: 11/27/19 09:40	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	2.3	ug/L	0.10	1		12/13/19 11:37	7439-92-1	
Sample: 15		Lab ID: 10501112009	Collected: 11/27/19 09:40	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.2	ug/L	0.10	1		12/13/19 11:39	7439-92-1	
Sample: 16		Lab ID: 10501112010	Collected: 11/27/19 09:40	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.8	ug/L	0.10	1		12/13/19 11:41	7439-92-1	
Sample: 17		Lab ID: 10501112011	Collected: 11/27/19 09:40	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.6	ug/L	0.10	1		12/13/19 11:43	7439-92-1	
Sample: 18		Lab ID: 10501112012	Collected: 11/27/19 09:40	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.5	ug/L	0.10	1		12/13/19 11:50	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Drinking Waters

Pace Project No.: 10501112

Sample: 38		Lab ID: 10501112013		Collected: 11/27/19 08:45		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 11:52	7439-92-1		
Sample: 39		Lab ID: 10501112014		Collected: 11/27/19 08:45		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 11:54	7439-92-1		
Sample: 40		Lab ID: 10501112015		Collected: 11/27/19 08:45		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 11:55	7439-92-1		
Sample: 41		Lab ID: 10501112016		Collected: 11/27/19 08:50		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 11:57	7439-92-1		
Sample: 42		Lab ID: 10501112017		Collected: 11/27/19 08:50		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 11:59	7439-92-1		
Sample: 43		Lab ID: 10501112018		Collected: 11/27/19 08:50		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 12:01	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Drinking Waters

Pace Project No.: 10501112

Sample: 44		Lab ID: 10501112019		Collected: 11/27/19 08:55		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 12:03	7439-92-1		
Sample: 45		Lab ID: 10501112020		Collected: 11/27/19 08:55		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 12:04	7439-92-1		
Sample: 46		Lab ID: 10501112021		Collected: 11/27/19 08:55		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 10:22	7439-92-1		
Sample: 29		Lab ID: 10501112022		Collected: 11/27/19 09:10		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 10:29	7439-92-1		
Sample: 30		Lab ID: 10501112023		Collected: 11/27/19 09:10		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 10:31	7439-92-1		
Sample: 31		Lab ID: 10501112024		Collected: 11/27/19 09:10		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 10:33	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Drinking Waters

Pace Project No.: 10501112

Sample: 32		Lab ID: 10501112025		Collected: 11/27/19 09:10		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 10:35	7439-92-1		
Sample: 33		Lab ID: 10501112026		Collected: 11/27/19 09:10		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 10:37	7439-92-1		
Sample: 34		Lab ID: 10501112027		Collected: 11/27/19 09:10		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 10:38	7439-92-1		
Sample: 35		Lab ID: 10501112028		Collected: 11/27/19 09:10		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 10:44	7439-92-1		
Sample: 36		Lab ID: 10501112029		Collected: 11/27/19 09:10		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 10:46	7439-92-1		
Sample: 37		Lab ID: 10501112030		Collected: 11/27/19 09:10		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	ND	ug/L	0.10	1		12/13/19 10:47	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Drinking Waters
Pace Project No.: 10501112

Sample: 1		Lab ID: 10501112031	Collected: 11/27/19 09:20	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.4	ug/L	0.10	1		12/13/19 10:49	7439-92-1	
Sample: 2		Lab ID: 10501112032	Collected: 11/27/19 09:20	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.6	ug/L	0.10	1		12/13/19 10:53	7439-92-1	
Sample: 3		Lab ID: 10501112033	Collected: 11/27/19 09:20	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.0	ug/L	0.10	1		12/13/19 10:55	7439-92-1	
Sample: 4		Lab ID: 10501112034	Collected: 11/27/19 09:20	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.53	ug/L	0.10	1		12/13/19 10:56	7439-92-1	
Sample: 5		Lab ID: 10501112035	Collected: 11/27/19 09:20	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.84	ug/L	0.10	1		12/13/19 10:58	7439-92-1	
Sample: 6		Lab ID: 10501112036	Collected: 11/27/19 09:20	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.74	ug/L	0.10	1		12/13/19 11:00	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Drinking Waters
Pace Project No.: 10501112

Sample: 19		Lab ID: 10501112037	Collected: 11/27/19 10:00	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.73	ug/L	0.10	1		12/13/19 11:06	7439-92-1	
Sample: 20		Lab ID: 10501112038	Collected: 11/27/19 10:00	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	ND	ug/L	0.10	1		12/13/19 11:07	7439-92-1	
Sample: 21		Lab ID: 10501112039	Collected: 11/27/19 10:00	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	ND	ug/L	0.10	1		12/13/19 11:09	7439-92-1	
Sample: 22		Lab ID: 10501112040	Collected: 11/27/19 10:00	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	ND	ug/L	0.10	1		12/13/19 11:11	7439-92-1	
Sample: 23		Lab ID: 10501112041	Collected: 11/27/19 10:00	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.11	ug/L	0.10	1		12/13/19 10:02	7439-92-1	
Sample: 24		Lab ID: 10501112042	Collected: 11/27/19 10:00	Received: 12/02/19 18:20	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.4	ug/L	0.10	1		12/13/19 10:10	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Drinking Waters

Pace Project No.: 10501112

Sample: 26		Lab ID: 10501112043		Collected: 11/27/19 10:00		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	0.17	ug/L	0.10	1		12/13/19 10:12	7439-92-1		

Sample: 27		Lab ID: 10501112044		Collected: 11/27/19 10:00		Received: 12/02/19 18:20		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8							
Lead	0.19	ug/L	0.10	1		12/13/19 10:14	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: Drinking Waters

Pace Project No.: 10501112

QC Batch:	648381	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	ICPMS Metals, Drinking Water
Associated Lab Samples:	10501112001, 10501112002, 10501112003, 10501112004, 10501112005, 10501112006, 10501112007, 10501112008, 10501112009, 10501112010, 10501112011, 10501112012, 10501112013, 10501112014, 10501112015, 10501112016, 10501112017, 10501112018, 10501112019, 10501112020		

METHOD BLANK:	3487857	Matrix:	Water
Associated Lab Samples:	10501112001, 10501112002, 10501112003, 10501112004, 10501112005, 10501112006, 10501112007, 10501112008, 10501112009, 10501112010, 10501112011, 10501112012, 10501112013, 10501112014, 10501112015, 10501112016, 10501112017, 10501112018, 10501112019, 10501112020		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	ND	0.10	12/13/19 11:13	

LABORATORY CONTROL SAMPLE: 3487858						
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	100	107	107	85-115	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE:												
3494013					3494014							
		10501112001	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits		Max RPD	Qual
Parameter	Units	Result								RPD		
Lead	ug/L	1.5	100	100	104	104	103	103	70-130	0	20	

MATRIX SPIKE SAMPLE:		3494015					
		10501112011	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	1.6	100	106	104	70-130	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: Drinking Waters

Pace Project No.: 10501112

QC Batch:	648384	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	ICPMS Metals, Drinking Water
Associated Lab Samples:	10501112021, 10501112022, 10501112023, 10501112024, 10501112025, 10501112026, 10501112027, 10501112028, 10501112029, 10501112030, 10501112031, 10501112032, 10501112033, 10501112034, 10501112035, 10501112036, 10501112037, 10501112038, 10501112039, 10501112040		

METHOD BLANK:	3487862	Matrix:	Water
Associated Lab Samples:	10501112021, 10501112022, 10501112023, 10501112024, 10501112025, 10501112026, 10501112027, 10501112028, 10501112029, 10501112030, 10501112031, 10501112032, 10501112033, 10501112034, 10501112035, 10501112036, 10501112037, 10501112038, 10501112039, 10501112040		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	ND	0.10	12/13/19 10:15	

LABORATORY CONTROL SAMPLE: 3487863						
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	100	106	106	85-115	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE:												
3494010					3494011							
Parameter	Units	10501112021 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Lead	ug/L	ND	100	100	106	103	106	103	70-130	3	20	

MATRIX SPIKE SAMPLE:		3494012					
		10501112031	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	1.4	100	103	101	70-130	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: Drinking Waters

Pace Project No.: 10501112

QC Batch: 648385 Analysis Method: EPA 200.8
QC Batch Method: EPA 200.8 Analysis Description: ICPMS Metals, Drinking Water
Associated Lab Samples: 10501112041, 10501112042, 10501112043, 10501112044

METHOD BLANK: 3487867 Matrix: Water
Associated Lab Samples: 10501112041, 10501112042, 10501112043, 10501112044

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	ND	0.10	12/13/19 09:55	

LABORATORY CONTROL SAMPLE: 3487868

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	100	106	106	85-115	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 3493926 3493927

Parameter	Units	10501112041 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Lead	ug/L	0.11	100	100	97.6	101	97	101	70-130	4	20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALIFIERS

Project: Drinking Waters

Pace Project No.: 10501112

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

LABORATORIES

PASI-M Pace Analytical Services - Minneapolis

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: Drinking Waters

Pace Project No.: 10501112

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
10501112001	7	EPA 200.8	648381		
10501112002	8	EPA 200.8	648381		
10501112003	9	EPA 200.8	648381		
10501112004	10	EPA 200.8	648381		
10501112005	11	EPA 200.8	648381		
10501112006	12	EPA 200.8	648381		
10501112007	13	EPA 200.8	648381		
10501112008	14	EPA 200.8	648381		
10501112009	15	EPA 200.8	648381		
10501112010	16	EPA 200.8	648381		
10501112011	17	EPA 200.8	648381		
10501112012	18	EPA 200.8	648381		
10501112013	38	EPA 200.8	648381		
10501112014	39	EPA 200.8	648381		
10501112015	40	EPA 200.8	648381		
10501112016	41	EPA 200.8	648381		
10501112017	42	EPA 200.8	648381		
10501112018	43	EPA 200.8	648381		
10501112019	44	EPA 200.8	648381		
10501112020	45	EPA 200.8	648381		
10501112021	46	EPA 200.8	648384		
10501112022	29	EPA 200.8	648384		
10501112023	30	EPA 200.8	648384		
10501112024	31	EPA 200.8	648384		
10501112025	32	EPA 200.8	648384		
10501112026	33	EPA 200.8	648384		
10501112027	34	EPA 200.8	648384		
10501112028	35	EPA 200.8	648384		
10501112029	36	EPA 200.8	648384		
10501112030	37	EPA 200.8	648384		
10501112031	1	EPA 200.8	648384		
10501112032	2	EPA 200.8	648384		
10501112033	3	EPA 200.8	648384		
10501112034	4	EPA 200.8	648384		
10501112035	5	EPA 200.8	648384		
10501112036	6	EPA 200.8	648384		
10501112037	19	EPA 200.8	648384		
10501112038	20	EPA 200.8	648384		
10501112039	21	EPA 200.8	648384		
10501112040	22	EPA 200.8	648384		
10501112041	23	EPA 200.8	648385		
10501112042	24	EPA 200.8	648385		
10501112043	26	EPA 200.8	648385		
10501112044	27	EPA 200.8	648385		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.



Section A Required Client Information:		Section B Required Project Information:		Section C Invoice Information:	
Company:	Report To:	Company Name:	Attention:	Page:	of
Address:	Copy To:	Address:		2294151	
NE Ham Lake, MN 55304		Purchase Order No.:		REGULATORY AGENCY	
Email To:		Project Name:		☐ NPDES ☐ GROUND WATER ☐ DRINKING WATER ☐ UST ☐ RCRA ☐ OTHER	
Phone: 763-754-6777		Project Number:		Site Location	
Requested Due Date/TAT:				STATE:	

ITEM #	Section D Required Client Information	Matrix Codes MATRIX / CODE	SAMPLE ID (A-Z, 0-9 / .)	SAMPLE TYPE (G=GRAB C=COMP)	COLLECTED		SAMPLE TEMP AT COLLECTION	# OF CONTAINERS	Preservatives	Analysis Test	Requested Analysis Filtered (Y/N)												Residual Chlorine (Y/N)	Pace Project No./ Lab I.D.
					COMPOSITE START	COMPOSITE END/GRAB																		
1		Drinking Water	38	D2G	11/27 8:00	11/27 8:00			Unpreserved	Analysis Test														013
2		Drinking Water	39	D2G	11/27 8:45	11/27 8:45			Unpreserved	Analysis Test														014
3		Drinking Water	40	D2G	11/27 8:45	11/27 8:45			Unpreserved	Analysis Test														015
4		Drinking Water	41	D2G	11/27 8:50	11/27 8:50			Unpreserved	Analysis Test														016
5		Drinking Water	42	D2G	11/27 8:50	11/27 8:50			Unpreserved	Analysis Test														017
6		Drinking Water	43	D2G	11/27 8:55	11/27 8:55			Unpreserved	Analysis Test														018
7		Drinking Water	44	D2G	11/27 8:55	11/27 8:55			Unpreserved	Analysis Test														019
8		Drinking Water	45	D2G	11/27 8:55	11/27 8:55			Unpreserved	Analysis Test														020
9		Drinking Water	46	D2G	11/27 9:10	11/27 9:10			Unpreserved	Analysis Test														021
10		Drinking Water	29	D2G	11/27 9:10	11/27 9:10			Unpreserved	Analysis Test														022
11		Drinking Water	30	D2G	11/27 9:10	11/27 9:10			Unpreserved	Analysis Test														023
12		Drinking Water	31	D2G	11/27 9:10	11/27 9:10			Unpreserved	Analysis Test														024

ADDITIONAL COMMENTS		RELINQUISHED BY / AFFILIATION	DATE	TIME	ACCEPTED BY / AFFILIATION	DATE	TIME	SAMPLE CONDITIONS	
Judy Abee 12.2.19 1320		Judy Abee	12.2.19	1400	Judy Abee	12.2.19	1400	Received on	Temp in °C
								Sealed Cooler	Custody
								Ice (Y/N)	(Y/N)
								Samples Intact	(Y/N)

SAMPLER NAME AND SIGNATURE	
PRINT Name of SAMPLER	DATE Signed (MM/DD/YYYY)
SIGNATURE OF SAMPLER	11/27/19

ORIGINAL

*Important Note: By signing this form you are accepting Pace's NET 30 day payment terms and agreeing to late charges of 1.5% per month for any invoices not paid within 30 days.

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

[illegible][illegible]

ORIGINAL		TEMP IN °C		Received on Ice (Y/N)	Custody Sealed Cooler (Y/N)	Samples Intact (Y/N)
SAMPLER NAME AND SIGNATURE						
PRINT Name of SAMPLER:	DATE Signed (MM/DD/YY):					
SIGNATURE of SAMPLER:						

	Document Name: Sample Condition Upon Receipt Form	Document Revised: 14Nov2019 Page 1 of 1
	Document No.: F-MN-L-213-rev.30	Pace Analytical Services - Minneapolis

Sample Condition Upon Receipt	Client Name: <u>DaVinci Academy</u>	Project #: WO#: 10501112
	Courier: ☐ Fed Ex ☐ UPS ☐ USPS ☐ Client ☒ Pace ☐ Speedee ☐ Commercial ☐ See Exceptions	PM: JDD Due Date: 12/17/19 CLIENT: DaVinci

Tracking Number: _____

Custody Seal on Cooler/Box Present? ☐ Yes ☒ No Seals Intact? ☐ Yes ☒ No Biological Tissue Frozen? ☐ Yes ☐ No ☒ N/A

Packing Material: ☒ Bubble Wrap ☐ Bubble Bags ☐ None ☐ Other: _____ Temp Blank? ☒ Yes ☐ No

Thermometer: ☐ T1(0461) ☐ T2(1336) ☐ T3(0459) Type of Ice: ☐ Wet ☐ Blue ☒ None ☐ Dry ☐ Melted
☐ T4(0254) ☒ T5(0489)

Note: Each West Virginia Sample must have temp taken (no temp blanks)

Temp should be above freezing to 6°C	Cooler Temp Read w/temp blank: _____ °C	Average Corrected Temp (no temp blank only): ☐ See Exceptions ☐ 1 Container
Correction Factor: <u>10.1</u>	Cooler Temp Corrected w/temp blank: _____ °C	

USDA Regulated Soil: (☒ N/A, water sample/Other: _____) Date/Initials of Person Examining Contents: CEG 12/2/19

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX or VA (check maps)? ☐ Yes ☐ No Did samples originate from a foreign source (internationally, including Hawaii and Puerto Rico)? ☐ Yes ☐ No

If Yes to either question, fill out a Regulated Soil Checklist (F-MN-Q-338) and include with SCUR/COC paperwork.

	COMMENTS:
Chain of Custody Present and Filled Out? ☒ Yes ☐ No	1.
Chain of Custody Relinquished? ☒ Yes ☐ No	2.
Sampler Name and/or Signature on COC? ☒ Yes ☐ No ☐ N/A	3.
Samples Arrived within Hold Time? ☒ Yes ☐ No	4.
Short Hold Time Analysis (<72 hr)? ☐ Yes ☒ No	5. ☐ Fecal Coliform ☐ HPC ☐ Total Coliform/E coli ☐ BOD/cBOD ☐ Hex Chrome ☐ Turbidity ☐ Nitrate ☐ Nitrite ☐ Orthophos ☐ Other
Rush Turn Around Time Requested? ☐ Yes ☒ No	6.
Sufficient Volume? ☒ Yes ☐ No	7.
Correct Containers Used? ☒ Yes ☐ No	8.
-Pace Containers Used? ☒ Yes ☐ No	9.
Containers Intact? ☒ Yes ☐ No	
Field Filtered Volume Received for Dissolved Tests? ☐ Yes ☐ No ☒ N/A	10. Is sediment visible in the dissolved container? ☐ Yes ☐ No
Is sufficient information available to reconcile the samples to the COC? ☒ Yes ☐ No	11. If no, write ID/ Date/Time on Container Below: <u>See Exception</u> <u>Sample 25 & 28 not received ID</u>
Matrix: ☒ Water ☐ Soil ☐ Oil ☐ Other	
All containers needing acid/base preservation have been checked? ☒ Yes ☐ No ☐ N/A	12. Sample # <u>1-461</u> ☐ NaOH ☒ HNO ₃ ☐ H ₂ SO ₄ ☐ Zinc Acetate
All containers needing preservation are found to be in compliance with EPA recommendation? ☐ Yes ☒ No ☐ N/A (HNO ₃ , H ₂ SO ₄ , <2pH, NaOH >9 Sulfide, NaOH >12 Cyanide)	Positive for Res. ☐ Yes ☐ No pH Paper Lot# <u>203619</u> See Exception ☒
Exceptions: VOA, Coliform, TOC/DOC Oil and Grease, DRO/8015 (water) and Dioxin/PFAS ☐ Yes ☐ No ☒ N/A	Res. Chlorine ☐ Yes ☐ No 0-6 Roll <u>203619</u> 0-6 Strip 0-14 Strip
Extra labels present on soil VOA or WIDRO containers? ☐ Yes ☐ No ☒ N/A	13. See Exception ☐
Headspace in VOA Vials (greater than 6mm)? ☐ Yes ☐ No ☒ N/A	
Trip Blank Present? ☐ Yes ☐ No ☒ N/A	14. Pace Trip Blank Lot # (if purchased): _____
Trip Blank Custody Seals Present? ☐ Yes ☐ No ☒ N/A	

CLIENT NOTIFICATION/RESOLUTION

Person Contacted: _____ Date/Time: _____

Comments/Resolution: _____

Field Data Required? ☐ Yes ☐ No

Project Manager Review: [Signature] Date: 12/3/19

Note: Whenever there is a discrepancy affecting North Carolina compliance samples, a copy of this form will be sent to the North Carolina DEHNR Certification Office (i.e. out of hold, incorrect preservative, out of temp, incorrect containers).

